

CERTIFICATE

of Conformity

Registration No.: AK 50625269 0001
Report No.: CN24GNJV 001
Holder: Additel Corporation
2900 Saturn Street #B Brea,
CA 92821
USA
Product: Laboratory Equipment
(*Digital Pressure Gauge*)

Type designation and serial number are listed on the next page(s)

The certificate of conformity refers to the above-mentioned product. This is to certify that the specimen is in conformity with the assessment requirement mentioned on the next page. This certificate does not imply assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity.

Certification Body

Date: 2024-04-17

Qiusheng Zhu



TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg

CERTIFICATE

of Conformity

Registration No.: AK 50625269 0001

Product: Laboratory Equipment
(Digital Pressure Gauge)

Tested according to: EN 61010-1:2010+A1
EN IEC 62311:2020
EN 301489-1 V 2.2.3:2019
EN 301489-17 V 3.2.4:2020
EN IEC 61326-1:2021
EN 300328 V 2.2.2:2019

Identification: Type Designation
Additel601, Additel601Ex, Additel601P,
Additel601PEX, Additel680A, Additel680AEx,
Additel680P, Additel680PEX(Additel)

Serial No.:

A003650452-001

Remark: 1) Refer to test report CN24GNJV 001 for details.

2) The test standards are for compliance with
Article 3.1a, 3.1b, 3.2 of Council Directive
2014/53/EU, referred to as the RED.



TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg