

# C E R T I F I C A T E

of Conformity

EC Council Directive 2014/53/EU

On Radio Equipment and Telecommunications Terminal Equipment

Registration No.: AT 50515202 0001

Report No.: CN21AAV8 001

Holder: Additel Corporation  
2900 Saturn Street #B Brea,  
CA 92821  
USA

Product: Laboratory Equipment  
(Laboratory Thermocouple Calibration Furnace)

Identification: Type Designation: ADT850 (Additel)  
To be continued.

Tested acc. to: EN 61010-1:2010+A1  
EN IEC 61010-2-010:2020  
EN 61010-2-030:2010  
EN IEC 62311:2020  
EN IEC 61000-3-2:2019

This certificate of conformity is based on an evaluation of a sample of the above mentioned product. This is to certify that the tested sample is in conformity with all provisions of Article 3 of Council Directive 2014/53/EU. This certificate does not imply assessment of the production and does not permit the use of a TÜV Rheinland mark of conformity. The holder of the certificate is authorized to use this certificate as part of the technical documentation and in combination with the EC Declaration of Conformity.

Date 16.09.2021



Certification Body

  
Mai Miao

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg

Phone: (+49/221)806-1371 Fax: (+49/221)806-3935 e-mail: cert-validity@de.tuv.com <http://www.tuv.com/safety>

CE The CE marking may only be used if all relevant and effective EC Directives are complied with. CE

# C E R T I F I C A T E

of Conformity

EC Council Directive 2014/53/EU

On Radio Equipment and Telecommunications Terminal Equipment



Registration No.: AT 50515202 0002

Report No.: CN21AAV8 001

Holder: Additel Corporation  
2900 Saturn Street #B Brea,  
CA 92821  
USA

Product: Laboratory Equipment  
(Laboratory Thermocouple Calibration Furnace)

Identification: Continued.  
Serial No. : Engineering sample  
Remark: Refer to test report CN21AAV8 001 for details.

Tested acc. to: EN 61000-3-3:2013+A1  
EN 61326-1:2013  
EN 300328 V 2.2.2:2019  
EN 301489-1 V 2.2.3:2019  
EN 301489-17 V 3.2.4:2020

This certificate of conformity is based on an evaluation of a sample of the above mentioned product. This is to certify that the tested sample is in conformity with all provisions of Article 3 of Council Directive 2014/53/EU. This certificate does not imply assessment of the production and does not permit the use of a TÜV Rheinland mark of conformity. The holder of the certificate is authorized to use this certificate as part of the technical documentation and in combination with the EC Declaration of Conformity.



Certification Body

Date 16.09.2021

  
Mai Miao

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg  
Phone:(+49/221)806-1371 Fax:(+49/221)806-3935 e-mail: cert-validity@de.tuv.com <http://www.tuv.com/safety>

CE The CE marking may only be used if all relevant and effective EC Directives are complied with. CE

Additel Corporation  
Mr. Hong Jun Gao

Date : 2021/09/16  
Our ref. : Zhaozhi 01  
Your ref.: H. J. G

2900 Saturn Street #B Brea,  
CA 92821  
USA

**Ref : AT RED Conformity**

Type of Equipment : Laboratory Thermocouple Calibration Furnace  
Model Designation : See Certificate  
Certificate No. : AT 50515202 0001  
Report No. : CN21AAV8 001

Dear Mr. Hong Jun Gao,


We herewith confirm that a sample of the above mentioned technical equipment has been tested and was found to be in accordance with the relevant requirements.

Enclosed please find your Certificate of Conformity.

We appreciate your kind support and would like to offer our assistance and continuous services in the future.

With kind regards,

Certification Body

Mai Miao 

CC: Beijing ConST Instruments

Enclosure

证书的详细资料请登陆[www.certipedia.com](http://www.certipedia.com)查阅,或拨打我司客服热线800 999 3668 / 400 883 1300咨询