

C E R T I F I C A T E

of

Conformity

Radio Equipment Regulations 2017

(UK SI 2017 No.1206)



Registration No.: AW 50515227 0001

Test Report No.: CN21WNWU 001

Holder: Additel Corporation
2900 Saturn Street #B Brea,
CA 92821
USA

Product: Laboratory Equipment
(Laboratory Thermocouple Calibration Furnace)

Identification: Type Designation: ADT850 (Additel)
To be continued.

Tested acc. to: BS EN 61010-1:2010+A1
BS EN IEC 61010-2-010:2020
BS EN 61010-2-030:2010
BS EN IEC 62311:2020
BS EN IEC 61000-3-2:2019

This certificate of conformity is based on an evaluation of a sample of the above mentioned product, technical report and documentation. This certificate does not imply the assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity.

Beijing, 17.09.2021



Certification Body


Mai Miao

TÜV Rheinland (China) Ltd. - No. 01/03B-08, Floor 7 and No. 01/04B-08, Floor 11, AVIC Building, No. 10B, Central Road East 3rd Ring Road, Chaoyang District, Beijing P.R. China 100022



The UKCA mark may be used, if all relevant UK Legislations and the UK Declaration of Conformity has been completed.



C E R T I F I C A T E

of

Conformity

Radio Equipment Regulations 2017

(UK SI 2017 No.1206)



Registration No.: AW 50515227 0002

Test Report No.: CN21WNWU 001

Holder: Additel Corporation
2900 Saturn Street #B Brea,
CA 92821
USA

Product: Laboratory Equipment
(Laboratory Thermocouple Calibration Furnace)

Identification: Continued.
Serial No. : Engineering sample
Remark: Refer to test report CN21WNWU 001 for details.

Tested acc. to: BS EN 61000-3-3:2013+A1
BS EN 61326-1:2013
EN 300328 V 2.2.2:2019
EN 301489-1 V 2.2.3:2019
EN 301489-17 V 3.2.4:2020

This certificate of conformity is based on an evaluation of a sample of the above mentioned product, technical report and documentation. This certificate does not imply the assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity.

Beijing, 17.09.2021



Certification Body


Mai Miao

TÜV Rheinland (China) Ltd. - No. 01/03B-08, Floor 7 and No. 01/04B-08,
Floor 11, AVIC Building, No. 10B, Central Road East 3rd Ring Road,
Chaoyang District, Beijing P.R. China 100022



The UKCA mark may be used, if all relevant UK Legislations and the UK Declaration of Conformity has been completed.



Additel Corporation
Mr. Hong Jun Gao

Date : 17.09.2021
Our ref. : Zhaozhi 01
Your ref.: H. J. G

2900 Saturn Street #B Brea,
CA 92821
USA

Ref : AW UK Radio Equipment Regulations 2017

Type of Equipment : Laboratory Thermocouple Calibration Furnace
Model Designation : See Certificate
Certificate No. : AW 50515227 0001
Report No. : CN21WNWU 001

Dear Mr. Hong Jun Gao,

We herewith confirm that a sample of the above mentioned technical equipment has been tested and was found to be in accordance with the relevant requirements.

Enclosed please find your Certificate of Conformity.

We appreciate your kind support and would like to offer our assistance and continuous services in the future.

With kind regards,

Certification Body

Mai Miao



CC: Beijing ConST Instruments

Enclosure

证书的详细资料请登陆www.certipedia.com查阅,或拨打我司客服热线800 999 3668 / 400 883 1300咨询